

~~SECRET~~

MS 2634

ENG H 60 - 818  
10 OCT 1960

ATTN :   
Chief, Operations Division, Medical Staff  
Chief, Engineering Staff, OC

25X1

Status of Radiological Survey and Radiation Detection Instruments  
 Quarterly Report

25X1

1. During July, August, and September the radiological survey and radiation detection equipments stored at  were tested to determine their operational status. Three instruments found to be inoperative were placed in operating condition.

25X1  
25X1

2. A total of four man-hours was expended in maintenance and repair during the last three months.

25X1

  
R&D/Lab/JWM/rkb (4 October 1960)

Distribution:

- Original and 1 - Addressee  
1 - Lab Subj. ✓  
1 - OC-E Chrono  
1 - Dev/s  
1 - Lab Pending

Please Return To  
Engineering Staff

DOC	11	REV DATE	5-12-80	BY	008632
ORIG COMP		ORIG CLASS	5	ORIG DATE	4
JUST	22	ORIG DATE	2010		

CONFIDENTIAL

~~SECRET~~

SECRET

END M 60-

Chief, Operations Division, Medical Staff

ATTN :

Chief, Engineering Staff, OC

Status of Radiological Survey and  
Radiation Detection Instruments

- Quarterly Report

1. During April, May, and June the radiological survey and radiation detection equipments were recalibrated and new batteries installed. With the exception of one instrument, those found inoperative were repaired and placed in operating condition.

2. One radiation detection meter, Radinc Type SUE S/M is inoperative because of damage to the ion chamber and other components. Because of the extensive damage, it is recommended that this meter be cannibalized for parts. Unless notified differently, salvage will begin in the next period and the meter will not be considered among the units tested each month.

3. A total of twenty-nine man-hours was expended during the last three months.

R&D/Lab/JWM/bao (20 Jun 60)

Distribution:

- Original & 1 - Addressee
- 1 - Lab Subject
- 1 - Lab Pending
- 1 - OC-E Chrono
- 1 - Lab Chrono

NOTE. Phone Call From 1000 hrs 14 July 1960  
stated HE RECEIVED AUTHORITY TO CANNIBALIZE

SECRET

CON

SECRET

449354

PLEASE RETURN TO R&D LAB

ATTN : Chief, Operations Division, Medical Staff

ENG M-60- 315

25X1

Chief, Engineering Staff, OC

Status of Radiological Survey and Radiation Detection  
Instruments [redacted] - Quarterly Report

Please Return To  
Engineering Staff

25X1

1. During January, February, and March the radiological survey and radiation detection instruments [redacted] were tested to determine their operational status. Twenty instruments were found to be inoperative. These instruments were repaired and placed in operating condition.

25X1

2. New batteries have been requested and should be received shortly. Upon receipt of the batteries calibration of the instruments will be started.

3. A total of seven man-hours was expended in maintenance and repair during the last three months.

*[Signature]*  
R&D/Lab/JWM/jcm (18 Apr 60) (Retyped from 6 Apr 60)

Distribution:

Original and 1 - Addressee

1 - Lab Subject

1 - Lab Pending

1 - OC-E Chrono

1 - Lab Chrono

PLEASE RETURN TO R&D LAB

25X1

CONFIDENTIAL

*M-46684*

Chief, Operations Division, Medical Staff  
ATTN :

ENG-M 9-1095

18 DEC 1959

25X1

Chief, Engineering Staff, OC

Status of Radiological Survey and Radiation Detection  
Instruments  Quarterly Report

Please Return To  
Engineering Staff

25X1


1. During October, November and December the radiological survey and radiation detection instruments  were tested to determine their operational status. A total of 18 instruments were found to be inoperative. These instruments were repaired and placed in operating condition.

25X1

2. Two instruments (Thyac type 389) are no longer serviceable because of unrepairable defective high voltage power supplies. These instruments have been cannibalized and other units have been repaired with the components taken from the cannibalized instruments.

3. A request for replacement components and batteries is being prepared for submission early in January 1960. Battery replacement and calibration will be initiated upon receipt of the replacement batteries.

4. The total man-hours expended the last quarter year: 7.  
The total man-hours expended in 1959: 45.

  
R&D/Lab/JWM/bao (15 Dec 59)



25X1

Distribution:

- Original & 1 - Addressee
- 1 - Lab Subject
- 1 - OC-E Chrono
- 1 - R&D Chrono
- 1 - Dev/s
- 1 - Lab Pending